

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011125.D
 Acq On : 27 Jul 2019 05:50
 Operator : JC/SP
 Sample : K3620-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D

Quant Time: Jul 29 01:53:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	194780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	317771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	305110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	141089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107229	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	5.50	113	97974	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	389521	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.13	95	137631	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
Target Compounds						
16) Acetone	2.45	43	10813	11.554	ug/l	95
20) Methylene Chloride	2.86	84	11124	5.609	ug/l	90
43) Isopropyl Acetate	6.45	43	10585	2.409	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011125.D
Acq On : 27 Jul 2019 05:50
Operator : JC/SP
Sample : K3620-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D

Quant Time: Jul 29 01:53:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

