

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023441.D
 Acq On : 26 Jul 2021 11:09
 Operator : JC/MD
 Sample : VX0726WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0726WBL01

Quant Time: Jul 27 03:26:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	177265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	301882	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	254026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	116614	52.737	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.480%	
35) Dibromofluoromethane	5.397	113	96000	50.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.720%	
50) Toluene-d8	8.653	98	356068	51.346	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.085	95	104412	44.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023441.D
Acq On : 26 Jul 2021 11:09
Operator : JC/MD
Sample : VX0726WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0726WBL01

Quant Time: Jul 27 03:26:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

