

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023447.D
 Acq On : 26 Jul 2021 13:39
 Operator : JC/MD
 Sample : M3145-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-10-PS

Quant Time: Jul 27 03:27:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	170250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	283387	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108703	51.185	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.380%
35) Dibromofluoromethane	5.397	113	89301	50.400	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.653	98	339565	52.162	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.320%
62) 4-Bromofluorobenzene	11.085	95	101464	45.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	2884	2.768	ug/l #	87
27) cis-1,2-Dichloroethene	4.501	96	8164	3.956	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023447.D
Acq On : 26 Jul 2021 13:39
Operator : JC/MD
Sample : M3145-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-10-PS

Quant Time: Jul 27 03:27:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

