

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023448.D
 Acq On : 26 Jul 2021 14:03
 Operator : JC/MD
 Sample : M3145-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-1-PDB

Quant Time: Jul 27 03:27:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	173729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	294816	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111916	51.643	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.280%		
35) Dibromofluoromethane	5.397	113	92729	50.305	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.620%		
50) Toluene-d8	8.653	98	344181	50.821	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.640%		
62) 4-Bromofluorobenzene	11.085	95	109420	47.514	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 95.020%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1935	1.820	ug/l	100
30) Chloroform	5.105	83	3276	0.951	ug/l	95
38) Carbon Tetrachloride	5.696	117	3984	1.469	ug/l #	94
44) Trichloroethene	7.135	130	1311	0.604	ug/l	82
64) Tetrachloroethene	9.281	164	530685	259.696	ug/l	97
90) Hexachloroethane	12.542	117	4022	5.324	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023448.D
Acq On : 26 Jul 2021 14:03
Operator : JC/MD
Sample : M3145-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-1-PDB

Quant Time: Jul 27 03:27:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

