

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023450.D
 Acq On : 26 Jul 2021 14:50
 Operator : JC/MD
 Sample : M3146-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-FD-02

Quant Time: Jul 27 03:27:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	170868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290264	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	258717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	111863	52.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.960%
35) Dibromofluoromethane	5.397	113	92788	51.127	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.260%
50) Toluene-d8	8.653	98	346275	51.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.860%
62) 4-Bromofluorobenzene	11.085	95	113177	49.916	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.840%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2946	2.818	ug/l #	83
27) cis-1,2-Dichloroethene	4.495	96	9148	4.417	ug/l	96
30) Chloroform	5.099	83	1687	0.498	ug/l #	72
64) Tetrachloroethene	9.281	164	1684	0.818	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023450.D
Acq On : 26 Jul 2021 14:50
Operator : JC/MD
Sample : M3146-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-FD-02

Quant Time: Jul 27 03:27:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

