

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023451.D
 Acq On : 26 Jul 2021 15:14
 Operator : JC/MD
 Sample : M3146-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-14S

Quant Time: Jul 27 03:27:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	169057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	286146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	254023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111057	52.663	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.320%			
35) Dibromofluoromethane	5.397	113	90856	50.783	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.560%			
50) Toluene-d8	8.653	98	342908	52.167	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.340%			
62) 4-Bromofluorobenzene	11.085	95	107713	48.190	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.380%			
Target Compounds						Qvalue
16) Acetone	2.392	43	3197	3.090	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	992	0.484	ug/l	94
30) Chloroform	5.099	83	2616	0.780	ug/l	97
64) Tetrachloroethene	9.281	164	89303	44.172	ug/l	99
90) Hexachloroethane	12.542	117	1318	3.069	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023451.D
Acq On : 26 Jul 2021 15:14
Operator : JC/MD
Sample : M3146-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-14S

Quant Time: Jul 27 03:27:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

