

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023452.D
 Acq On : 26 Jul 2021 15:37
 Operator : JC/MD
 Sample : M3146-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-1D(ONSITE)

Quant Time: Jul 27 03:28:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	159563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	269453	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104339	52.421	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.840%	
35) Dibromofluoromethane	5.397	113	85202	50.573	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.140%	
50) Toluene-d8	8.653	98	324211	52.378	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.760%	
62) 4-Bromofluorobenzene	11.085	95	97254	46.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.420%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2350	2.407	ug/l	99
44) Trichloroethene	7.129	130	3239	1.631	ug/l	97
64) Tetrachloroethene	9.281	164	2257	1.213	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023452.D
Acq On : 26 Jul 2021 15:37
Operator : JC/MD
Sample : M3146-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-1D(ONSITE)

Quant Time: Jul 27 03:28:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

