

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023453.D
 Acq On : 26 Jul 2021 16:01
 Operator : JC/MD
 Sample : M3146-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-15D

Quant Time: Jul 27 03:28:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	166434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	278906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107296	51.681	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.360%
35) Dibromofluoromethane	5.397	113	88992	51.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	8.653	98	339895	53.051	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.100%
62) 4-Bromofluorobenzene	11.085	95	110368	50.660	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.320%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2871	2.819	ug/l #	87
30) Chloroform	5.105	83	2765	0.837	ug/l #	81
44) Trichloroethene	7.129	130	609	0.296	ug/l	72
64) Tetrachloroethene	9.275	164	78569	38.839	ug/l	94
90) Hexachloroethane	12.542	117	1369	3.051	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023453.D
Acq On : 26 Jul 2021 16:01
Operator : JC/MD
Sample : M3146-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-15D

Quant Time: Jul 27 03:28:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

