

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023454.D
 Acq On : 26 Jul 2021 16:24
 Operator : JC/MD
 Sample : M3146-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-10-PD

Quant Time: Jul 27 03:28:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	158484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	267769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103388	52.297	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.600%
35) Dibromofluoromethane	5.397	113	85772	51.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.460%
50) Toluene-d8	8.653	98	321376	52.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.500%
62) 4-Bromofluorobenzene	11.085	95	96583	46.176	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.360%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1119	1.154	ug/l #	69
20) Methylene Chloride	2.800	84	1212	0.643	ug/l	93
64) Tetrachloroethene	9.275	164	653	0.348	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023454.D
Acq On : 26 Jul 2021 16:24
Operator : JC/MD
Sample : M3146-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-10-PD

Quant Time: Jul 27 03:28:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

