

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023456.D
 Acq On : 26 Jul 2021 17:12
 Operator : JC/MD
 Sample : M3146-03 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-6-PS

Quant Time: Jul 27 03:28:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167726	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	281305	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109686	52.425	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.860%	
35) Dibromofluoromethane	5.397	113	89425	50.843	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.680%	
50) Toluene-d8	8.653	98	336016	51.998	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.000%	
62) 4-Bromofluorobenzene	11.085	95	108642	49.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.880%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.501	96	1259	0.619	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023456.D
Acq On : 26 Jul 2021 17:12
Operator : JC/MD
Sample : M3146-03 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-6-PS

Quant Time: Jul 27 03:28:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

