

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023457.D
 Acq On : 26 Jul 2021 17:35
 Operator : JC/MD
 Sample : M3146-04 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-9-PD

Quant Time: Jul 27 03:28:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	163452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81907	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106551	52.259	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.520%
35) Dibromofluoromethane	5.397	113	88009	51.147	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.300%
50) Toluene-d8	8.653	98	328612	51.979	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.960%
62) 4-Bromofluorobenzene	11.085	95	99153	46.124	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023457.D
Acq On : 26 Jul 2021 17:35
Operator : JC/MD
Sample : M3146-04 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-9-PD

Quant Time: Jul 27 03:28:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

