

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
 Data File : VX023458.D
 Acq On : 26 Jul 2021 17:59
 Operator : JC/MD
 Sample : M3146-09 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-8-PS

Quant Time: Jul 27 03:29:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	162543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105998	52.278	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.560%	
35) Dibromofluoromethane	5.397	113	87722	51.015	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.020%	
50) Toluene-d8	8.653	98	331661	52.498	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.000%	
62) 4-Bromofluorobenzene	11.085	95	99535	46.333	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.660%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	1179	0.598	ug/l	Qvalue 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072621\
Data File : VX023458.D
Acq On : 26 Jul 2021 17:59
Operator : JC/MD
Sample : M3146-09 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-8-PS

Quant Time: Jul 27 03:29:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

