

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030211.D
 Acq On : 26 Jul 2022 15:40
 Operator : JC/MD
 Sample : N3876-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-07-07192022

Quant Time: Jul 26 17:22:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	131950	53.111	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.220%
35) Dibromofluoromethane	5.391	113	103655	48.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.860%
50) Toluene-d8	8.653	98	386881	47.155	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.320%
62) 4-Bromofluorobenzene	11.085	95	132688	43.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.160%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	11790	9.269	ug/l	88
25) 2-Butanone	4.574	43	3753	1.756	ug/l	92
64) Tetrachloroethene	9.275	164	1861	0.786	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030211.D
Acq On : 26 Jul 2022 15:40
Operator : JC/MD
Sample : N3876-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-07-07192022

Quant Time: Jul 26 17:22:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

