

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030213.D
 Acq On : 26 Jul 2022 16:26
 Operator : JC/MD
 Sample : N3876-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-08-07202022

Quant Time: Jul 26 18:13:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	144842	55.607	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.220%
35) Dibromofluoromethane	5.391	113	114434	51.356	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	8.653	98	418852	48.539	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.080%
62) 4-Bromofluorobenzene	11.085	95	142271	44.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.840%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	21873	16.402	ug/l	93
25) 2-Butanone	4.592	43	3632	1.621	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	1937	0.236	ug/l #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030213.D
Acq On : 26 Jul 2022 16:26
Operator : JC/MD
Sample : N3876-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-08-072022

Quant Time: Jul 26 18:13:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

