

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030215.D
 Acq On : 26 Jul 2022 17:13
 Operator : JC/MD
 Sample : N3876-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-07212022

Quant Time: Jul 26 18:14:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233620	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	133380	49.961	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.920%
35) Dibromofluoromethane	5.391	113	105652	47.308	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.620%
50) Toluene-d8	8.653	98	389768	45.066	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.140%
62) 4-Bromofluorobenzene	11.085	95	135158	42.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.220%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3166	2.316	ug/l #	79
64) Tetrachloroethene	9.275	164	5636	2.245	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030215.D
Acq On : 26 Jul 2022 17:13
Operator : JC/MD
Sample : N3876-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-07212022

Quant Time: Jul 26 18:14:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

