

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030216.D
 Acq On : 26 Jul 2022 17:37
 Operator : JC/MD
 Sample : N3876-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-06-84

Quant Time: Jul 26 18:14:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230167	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	139921	54.455	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.920%	
35) Dibromofluoromethane	5.391	113	111884	50.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.700%	
50) Toluene-d8	8.653	98	410209	48.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.280%	
62) 4-Bromofluorobenzene	11.085	95	141664	44.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.600%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1598	1.215	ug/l	94
44) Trichloroethene	7.129	130	3378	1.241	ug/l	94
64) Tetrachloroethene	9.275	164	267780	107.068	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030216.D
Acq On : 26 Jul 2022 17:37
Operator : JC/MD
Sample : N3876-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-06-84

Quant Time: Jul 26 18:14:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

