

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030217.D
 Acq On : 26 Jul 2022 18:00
 Operator : JC/MD
 Sample : N3876-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-11-84

Quant Time: Jul 26 18:22:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232642	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126702	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130283	48.975	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.940%
35) Dibromofluoromethane	5.391	113	102707	46.182	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.360%
50) Toluene-d8	8.653	98	375981	43.655	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.320%
62) 4-Bromofluorobenzene	11.085	95	130548	40.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.680%#
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3398	2.495	ug/l	98
44) Trichloroethene	7.135	130	2565	0.932	ug/l	83
64) Tetrachloroethene	9.281	164	2031	0.802	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030217.D
Acq On : 26 Jul 2022 18:00
Operator : JC/MD
Sample : N3876-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-11-84

Quant Time: Jul 26 18:22:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

