

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030218.D
 Acq On : 26 Jul 2022 18:24
 Operator : JC/MD
 Sample : N3876-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-07-84

Quant Time: Jul 26 19:07:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	128621	47.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.160%
35) Dibromofluoromethane	5.397	113	99515	44.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.080%
50) Toluene-d8	8.653	98	372955	42.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.240%#
62) 4-Bromofluorobenzene	11.086	95	127240	39.177	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.360%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4124	2.980	ug/l #	88
44) Trichloroethene	7.129	130	2211	0.791	ug/l	95
64) Tetrachloroethene	9.275	164	1439	0.566	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030218.D
Acq On : 26 Jul 2022 18:24
Operator : JC/MD
Sample : N3876-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-07-84

Quant Time: Jul 26 19:07:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

