

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030204.D
 Acq On : 26 Jul 2022 10:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 17:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	-0.01
2 T	Dichlorodifluoromethane	0.820	0.916	-11.7	108	0.00
3 P	Chloromethane	0.811	0.773	4.7	98	0.00
4 C	Vinyl Chloride	0.819	0.835	-2.0#	106	0.00
5 T	Bromomethane	0.355	0.304	14.4	104	0.00
6 T	Chloroethane	0.470	0.484	-3.0	109	0.00
7 T	Trichlorofluoromethane	1.079	1.163	-7.8	111	0.00
8 T	Diethyl Ether	0.422	0.427	-1.2	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.799	0.851	-6.5	113	0.00
10 T	Methyl Iodide	0.696	0.808	-16.1	97	0.00
11 T	Tert butyl alcohol	0.245	0.251	-2.4	106	0.00
12 CM	1,1-Dichloroethene	0.781	0.816	-4.5#	108	0.00
13 T	Acrolein	0.060	0.058	3.3	109	0.00
14 T	Allyl chloride	1.612	1.664	-3.2	110	0.00
15 T	Acrylonitrile	0.606	0.619	-2.1	103	0.00
16 T	Acetone	0.515	0.516	-0.2	105	0.00
17 T	Carbon Disulfide	2.108	2.128	-0.9	106	0.00
18 T	Methyl Acetate	1.522	1.522	0.0	108	0.00
19 T	Methyl tert-butyl Ether	2.676	2.851	-6.5	107	0.00
20 T	Methylene Chloride	1.184	0.976	17.6	106	0.00
21 T	trans-1,2-Dichloroethene	0.862	0.883	-2.4	107	0.00
22 T	Diisopropyl ether	3.050	3.355	-10.0	109	0.00
23 T	Vinyl Acetate	2.645	2.925	-10.6	107	0.00
24 P	1,1-Dichloroethane	1.620	1.681	-3.8	108	0.00
25 T	2-Butanone	0.865	0.897	-3.7	105	0.00
26 T	2,2-Dichloropropane	1.172	1.276	-8.9	119	0.00
27 T	cis-1,2-Dichloroethene	1.014	1.054	-3.9	108	0.00
28 T	Bromochloromethane	0.685	0.730	-6.6	114	0.00
29 T	Tetrahydrofuran	0.582	0.605	-4.0	106	0.00
30 C	Chloroform	1.604	1.684	-5.0#	108	0.00
31 T	Cyclohexane	1.459	1.597	-9.5	112	0.00
32 T	1,1,1-Trichloroethane	1.304	1.405	-7.7	110	0.00
33 S	1,2-Dichloroethane-d4	1.005	1.049	-4.4	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.478	0.502	-5.0	104	0.00
36 T	1,1-Dichloropropene	0.688	0.742	-7.8	109	0.00
37 T	Ethyl Acetate	0.974	0.980	-0.6	103	0.00
38 T	Carbon Tetrachloride	0.642	0.700	-9.0	110	0.00
39 T	Methylcyclohexane	0.888	0.952	-7.2	113	0.00
40 TM	Benzene	2.164	2.299	-6.2	107	0.00
41 T	Methacrylonitrile	0.522	0.559	-7.1	106	0.00
42 TM	1,2-Dichloroethane	0.748	0.778	-4.0	105	0.00
43 T	Isopropyl Acetate	1.428	1.540	-7.8	107	0.00
44 TM	Trichloroethene	0.591	0.596	-0.8	107	0.00
45 C	1,2-Dichloropropane	0.577	0.614	-6.4#	109	0.00
46 T	Dibromomethane	0.393	0.407	-3.6	108	0.00
47 T	Bromodichloromethane	0.714	0.769	-7.7	109	0.00
48 T	Methyl methacrylate	0.703	0.765	-8.8	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
 Data File : VX030204.D
 Acq On : 26 Jul 2022 10:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 17:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.016	0.017	-6.3	108	0.00
50 S	Toluene-d8	1.851	1.937	-4.6	102	0.00
51 T	4-Methyl-2-Pentanone	0.969	1.037	-7.0	106	0.00
52 CM	Toluene	1.320	1.419	-7.5#	106	0.00
53 T	t-1,3-Dichloropropene	0.769	0.854	-11.1	109	0.00
54 T	cis-1,3-Dichloropropene	0.838	0.924	-10.3	109	0.00
55 T	1,1,2-Trichloroethane	0.562	0.584	-3.9	105	0.00
56 T	Ethyl methacrylate	0.889	0.982	-10.5	106	0.00
57 T	1,3-Dichloropropane	0.928	0.977	-5.3	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.441	0.540	-22.4	113	0.00
59 T	2-Hexanone	0.765	0.816	-6.7	104	0.00
60 T	Dibromochloromethane	0.520	0.574	-10.4	107	0.00
61 T	1,2-Dibromoethane	0.587	0.610	-3.9	105	0.00
62 S	4-Bromofluorobenzene	0.687	0.742	-8.0	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.538	0.561	-4.3	106	0.00
65 PM	Chlorobenzene	1.358	1.431	-5.4	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.462	0.511	-10.6	106	0.00
67 C	Ethyl Benzene	2.659	2.650	0.3#	107	0.00
68 T	m/p-Xylenes	0.920	1.012	-10.0	106	0.00
69 T	o-Xylene	0.907	1.001	-10.4	106	0.00
70 T	Styrene	1.490	1.677	-12.6	105	0.00
71 P	Bromoform	0.339	0.394	-16.2	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.903	4.067	-4.2	108	0.00
74 T	N-amyl acetate	1.897	2.049	-8.0	107	0.00
75 P	1,1,2,2-Tetrachloroethane	1.429	1.401	2.0	105	0.00
76 T	1,2,3-Trichloropropane	1.318	1.379	-4.6	111	0.00
77 T	Bromobenzene	0.916	0.901	1.6	102	0.00
78 T	n-propylbenzene	4.659	4.953	-6.3	108	0.00
79 T	2-Chlorotoluene	2.942	2.911	1.1	107	0.00
80 T	1,3,5-Trimethylbenzene	3.268	3.418	-4.6	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.439	0.475	-8.2	106	0.00
82 T	4-Chlorotoluene	3.214	3.366	-4.7	108	0.00
83 T	tert-Butylbenzene	3.029	3.197	-5.5	108	0.00
84 T	1,2,4-Trimethylbenzene	3.261	3.473	-6.5	106	0.00
85 T	sec-Butylbenzene	4.146	4.444	-7.2	108	0.00
86 T	p-Isopropyltoluene	3.317	3.555	-7.2	108	0.00
87 T	1,3-Dichlorobenzene	1.782	1.793	-0.6	107	0.00
88 T	1,4-Dichlorobenzene	1.814	1.796	1.0	106	0.00
89 T	n-Butylbenzene	3.088	3.487	-12.9	112	0.00
90 T	Hexachloroethane	0.571	0.608	-6.5	110	0.00
91 T	1,2-Dichlorobenzene	1.749	1.783	-1.9	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.343	0.351	-2.3	109	0.00
93 T	1,2,4-Trichlorobenzene	1.014	1.125	-10.9	115	0.00
94 T	Hexachlorobutadiene	0.407	0.433	-6.4	113	0.00
95 T	Naphthalene	3.962	4.372	-10.3	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072622\
Data File : VX030204.D
Acq On : 26 Jul 2022 10:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 26 17:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.040	1.137	-9.3	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6