

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072624\
 Data File : VX042562.D
 Acq On : 25 Jul 2024 12:44
 Operator : JC/MD
 Sample : P3355-01DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-BLVD(JUNEDL)

Quant Time: Jul 26 01:35:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X071924W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 19 11:30:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	20539	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	113269	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	103080	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	45139	30.681	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.267%
60) 4-Bromofluorobenzene	11.079	95	44334	27.643	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.133%
63) Toluene-d8	8.647	98	133437	29.799	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.333%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	6771	30.797	ug/l	99
16) Carbon Disulfide	2.502	76	1507	0.637	ug/l	99
18) Methylene Chloride	2.788	84	2030	1.484	ug/l	94
30) 2-Butanone	4.587	43	3530	3.580	ug/l #	90
58) 4-Methyl-2-Pentanone	8.580	43	1300	0.716	ug/l	96
62) Toluene	8.720	91	179982	34.924	ug/l	100
84) 1,4-Dichlorobenzene	12.043	146	966	0.333	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072624\
Data File : VX042562.D
Acq On : 25 Jul 2024 12:44
Operator : JC/MD
Sample : P3355-01DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-BLVD(JUNEDL

Quant Time: Jul 26 01:35:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X071924W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 19 11:30:14 2024
Response via : Initial Calibration

