

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072718\
 Data File : VX003719.D
 Acq On : 27 Jul 2018 16:44
 Operator : JC/SP
 Sample : VX0727WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0727WBL01

Quant Time: Jul 28 02:35:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X072718W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 27 16:17:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	22052	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	100922	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	94164	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	46225	28.55	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	95.17%		
60) 4-Bromofluorobenzene	11.14	95	39838	26.28	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	87.60%		
63) Toluene-d8	8.71	98	121856	28.73	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	95.77%		

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072718\
Data File : VX003719.D
Acq On : 27 Jul 2018 16:44
Operator : JC/SP
Sample : VX0727WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0727WBL01

Quant Time: Jul 28 02:35:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X072718W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 27 16:17:43 2018
Response via : Initial Calibration

