

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072720\
 Data File : VX017635.D
 Acq On : 27 Jul 2020 22:21
 Operator : JC/SP
 Sample : L3421-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-3D

Quant Time: Jul 28 02:19:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	490780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	767107	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	773499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	444800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	314285	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
35) Dibromofluoromethane	5.46	113	262747	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
50) Toluene-d8	8.70	98	898815	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.12	95	409269	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
Target Compounds						
16) Acetone	2.42	43	8966	3.231	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072720\
Data File : VX017635.D
Acq On : 27 Jul 2020 22:21
Operator : JC/SP
Sample : L3421-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-3D

Quant Time: Jul 28 02:19:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

