

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072720\
 Data File : VX017633.D
 Acq On : 27 Jul 2020 21:35
 Operator : JC/SP
 Sample : L3421-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-6

Quant Time: Jul 28 02:15:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	497901	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	848924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	889646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	481446	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	340103	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
35) Dibromofluoromethane	5.47	113	286261	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	8.70	98	1051382	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	11.12	95	464494	56.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.34%	
Target Compounds						
16) Acetone	2.42	43	17113	6.079	ug/l	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072720\
Data File : VX017633.D
Acq On : 27 Jul 2020 21:35
Operator : JC/SP
Sample : L3421-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-6

Quant Time: Jul 28 02:15:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 746

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35

Instrument :

MSVOA_X

ClientSampleId :


SY-6

Tot Ion:168 Resp: 497901

Ion Ratio Lower Upper

168 100

99 55.5 43.4 65.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.64

200000



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 6.079 ug/l

RT: 2.42 min Scan# 219

Delta R.T. -0.00 min

Lab File: VX017633.D


Acq: 27 Jul 2020 21:35

Tgt Ion: 43 Resp: 17113

Ion Ratio Lower Upper

43 100

58 32.0 24.2 36.4



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.42

10000

8000


6000


4000

2000

0

Time-->





#33

1,2-Dichloroethane-d4

Concen: 53.176 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35

Instrument :

MSVOA_X

ClientSampled :


SY-6

Tot Ion: 65 Resp: 340103

Ion Ratio Lower Upper

65 100

67 50.6 0.0 102.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.03



150000

100000

50000

0

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35


Tgt Ion: 114 Resp: 848924

Ion Ratio Lower Upper

114 100

63 20.3 0.0 40.6

88 16.0 0.0 33.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0

6.83

400000


300000


200000

100000

0

Time-->





#35

Dibromofluoromethane

Concen: 50.419 ug/l

RT: 5.47 min Scan# 719

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35

Instrument :

MSVOA_X

ClientSampled :

SY-6


Tot Ion:113 Resp: 286261

Ion Ratio Lower Upper

113 100

111 101.4 80.7 121.1




192 20.7 16.9 25.3



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#50

Toluene-d8

Concen: 51.332 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX017633.D


Acq: 27 Jul 2020 21:35

Tgt Ion: 98 Resp: 1051382

Ion Ratio Lower Upper



98 100


100 63.2 52.3 78.5



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V





#62

4-Bromofluorobenzene

Concen: 56.667 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35

Instrument :

MSVOA_X

ClientSampleId :

SY-6


Tot Ion: 95 Resp: 464494

Ion Ratio Lower Upper

95 100

174 85.5 0.0 169.0




176 81.3 0.0 161.0



Abundance Ion 94.90 (94.60 to 95.60): VX017484.D (-1639) (-)

Ion 173.80 (173.50 to 174.50): VX017484.D (-1639) (-)

Ion 175.80 (175.50 to 176.50): VX017484.D (-1639) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35


Tgt Ion: 117 Resp: 889646

Ion Ratio Lower Upper

117 100

82 53.8 44.0 66.0



119 30.7 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): VX017484.D (-1467) (-)

Ion 82.00 (81.70 to 82.70): VX017484.D (-1467) (-)

Ion 118.90 (118.60 to 119.60): VX017484.D (-1467) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX017633.D

Acq: 27 Jul 2020 21:35

Instrument :

MSVOA_X

ClientSampleId :

SY-6


Tot Ion:152 Resp: 481446

Ion Ratio Lower Upper

152 100

115 59.8 65.5 196.6#

150 157.3 0.0 412.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

