

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072721\
 Data File : VX023474.D
 Acq On : 27 Jul 2021 14:02
 Operator : JC/MD
 Sample : M3146-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-6-PS

Quant Time: Jul 28 00:51:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	160156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	270165	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103552	51.833	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.660%			
35) Dibromofluoromethane	5.398	113	86575	51.252	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.500%			
50) Toluene-d8	8.653	98	323476	52.122	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.240%			
62) 4-Bromofluorobenzene	11.086	95	102839	48.731	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.460%			
Target Compounds						Qvalue
16) Acetone	2.386	43	2537	2.589	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	10401	5.357	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072721\
Data File : VX023474.D
Acq On : 27 Jul 2021 14:02
Operator : JC/MD
Sample : M3146-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-6-PS

Quant Time: Jul 28 00:51:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

