

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030227.D
 Acq On : 27 Jul 2022 12:44
 Operator : JC/MD
 Sample : N3876-02DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-06-84DL

Quant Time: Jul 28 05:49:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246944	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154631	54.042	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.080%
35) Dibromofluoromethane	5.391	113	122587	51.929	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.860%
50) Toluene-d8	8.653	98	447600	48.961	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.920%
62) 4-Bromofluorobenzene	11.085	95	156333	46.078	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.160%
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	1939	0.664	ug/l	84
64) Tetrachloroethene	9.275	164	110137	41.459	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030227.D
Acq On : 27 Jul 2022 12:44
Operator : JC/MD
Sample : N3876-02DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-06-84DL

Quant Time: Jul 28 05:49:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

