

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030228.D
 Acq On : 27 Jul 2022 13:54
 Operator : JC/MD
 Sample : N3876-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-11-84

Quant Time: Jul 28 05:50:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233766	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	145593	54.542	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.080%
35) Dibromofluoromethane	5.391	113	113743	50.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.800%
50) Toluene-d8	8.653	98	420988	48.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.085	95	144105	44.868	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.740%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3937	2.881	ug/l	91
44) Trichloroethene	7.141	130	2324	0.841	ug/l	100
64) Tetrachloroethene	9.269	164	1087	0.420	ug/l #	71
78) n-propylbenzene	11.305	91	5579	0.462	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030228.D
Acq On : 27 Jul 2022 13:54
Operator : JC/MD
Sample : N3876-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-11-84

Quant Time: Jul 28 05:50:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

