

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030229.D
 Acq On : 27 Jul 2022 14:17
 Operator : JC/MD
 Sample : N3876-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-07-84

Quant Time: Jul 28 05:50:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140040	52.611	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.220%
35) Dibromofluoromethane	5.391	113	109432	49.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.840%
50) Toluene-d8	8.653	98	403686	47.078	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.085	95	139338	43.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.560%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	4239	3.111	ug/l #	88
44) Trichloroethene	7.135	130	2120	0.774	ug/l	98
64) Tetrachloroethene	9.281	164	949	0.376	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030229.D
Acq On : 27 Jul 2022 14:17
Operator : JC/MD
Sample : N3876-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-07-84

Quant Time: Jul 28 05:50:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

