

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030230.D
 Acq On : 27 Jul 2022 14:41
 Operator : JC/MD
 Sample : N3876-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-10-84

Quant Time: Jul 28 05:50:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	145171	53.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.280%	
35) Dibromofluoromethane	5.391	113	115084	51.643	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.280%	
50) Toluene-d8	8.653	98	418897	48.539	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.080%	
62) 4-Bromofluorobenzene	11.085	95	145291	45.363	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.720%	
Target Compounds						Qvalue
16) Acetone	2.386	43	3829	2.763	ug/l	92
44) Trichloroethene	7.129	130	1025	0.372	ug/l	86
64) Tetrachloroethene	9.281	164	792	0.310	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030230.D
Acq On : 27 Jul 2022 14:41
Operator : JC/MD
Sample : N3876-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-10-84

Quant Time: Jul 28 05:50:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

