

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030231.D
 Acq On : 27 Jul 2022 15:04
 Operator : JC/MD
 Sample : N3765-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-04-07132022

Quant Time: Jul 28 05:50:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	144148	56.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.580%
35) Dibromofluoromethane	5.391	113	114896	52.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.100%
50) Toluene-d8	8.653	98	415382	48.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.085	95	143411	45.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	13973	10.658	ug/l	90
17) Carbon Disulfide	2.514	76	37724	7.022	ug/l	95
25) 2-Butanone	4.587	43	4378	1.987	ug/l #	89
64) Tetrachloroethene	9.275	164	2959	1.195	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030231.D
Acq On : 27 Jul 2022 15:04
Operator : JC/MD
Sample : N3765-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-04-07132022

Quant Time: Jul 28 05:50:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

