

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030234.D
 Acq On : 27 Jul 2022 16:14
 Operator : JC/MD
 Sample : N3765-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-07182022

Quant Time: Jul 28 05:51:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234650	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143990	53.430	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.860%	
35) Dibromofluoromethane	5.391	113	114338	50.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.940%	
50) Toluene-d8	8.653	98	417025	48.006	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.020%	
62) 4-Bromofluorobenzene	11.085	95	146976	45.589	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.180%	
Target Compounds						Qvalue
64) Tetrachloroethene	9.275	164	267850	103.998	ug/l	97
90) Hexachloroethane	12.542	117	626	0.418	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030234.D
Acq On : 27 Jul 2022 16:14
Operator : JC/MD
Sample : N3765-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-07182022

Quant Time: Jul 28 05:51:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

