

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030242.D
 Acq On : 27 Jul 2022 19:22
 Operator : JC/MD
 Sample : N3884-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-07-25

Quant Time: Jul 28 05:53:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223073	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	131581	52.563	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.120%
35) Dibromofluoromethane	5.391	113	103389	48.483	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.960%
50) Toluene-d8	8.653	98	371585	44.995	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.000%
62) 4-Bromofluorobenzene	11.085	95	128930	42.067	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.140%
Target Compounds						
16) Acetone	2.380	43	3049	2.379	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030242.D
Acq On : 27 Jul 2022 19:22
Operator : JC/MD
Sample : N3884-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-07-25

Quant Time: Jul 28 05:53:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

