

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030254.D
 Acq On : 28 Jul 2022 00:50
 Operator : JC/MD
 Sample : PB146425ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB146425ZHE#01

Quant Time: Jul 28 05:56:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	110409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	206812	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205616	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134914	60.795	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.600%	
35) Dibromofluoromethane	5.391	113	103715	52.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.920%	
50) Toluene-d8	8.653	98	368344	48.110	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.220%	
62) 4-Bromofluorobenzene	11.085	95	129686	45.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.280%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	37071	32.629	ug/l	99
18) Methyl Acetate	2.715	43	5459	1.624	ug/l	96
20) Methylene Chloride	2.788	84	16518	6.264	ug/l	98
25) 2-Butanone	4.568	43	10959	5.740	ug/l	94
37) Ethyl Acetate	4.721	43	46032	11.425	ug/l	98
43) Isopropyl Acetate	6.348	43	192510	32.594	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.085	75	65483	67.519	ug/l #	13

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030254.D
Acq On : 28 Jul 2022 00:50
Operator : JC/MD
Sample : PB146425ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB146425ZHE#01

Quant Time: Jul 28 05:56:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

