

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
 Data File : VX030255.D
 Acq On : 28 Jul 2022 01:13
 Operator : JC/MD
 Sample : PB146425ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB146425ZHE#02

Quant Time: Jul 28 05:56:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	226927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	144888	58.036	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.080%
35) Dibromofluoromethane	5.397	113	116178	53.555	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.120%
50) Toluene-d8	8.653	98	427149	50.845	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.700%
62) 4-Bromofluorobenzene	11.085	95	151352	48.544	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	37136	29.054	ug/l	93
18) Methyl Acetate	2.703	43	6108	1.616	ug/l	98
20) Methylene Chloride	2.788	84	16035	5.039	ug/l	89
37) Ethyl Acetate	4.721	43	46907	10.610	ug/l	97
43) Isopropyl Acetate	6.349	43	191286	29.516	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072722\
Data File : VX030255.D
Acq On : 28 Jul 2022 01:13
Operator : JC/MD
Sample : PB146425ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB146425ZHE#02

Quant Time: Jul 28 05:56:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

