

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072722\  
 Data File : VX030261.D  
 Acq On : 28 Jul 2022 03:33  
 Operator : JC/MD  
 Sample : N3884-02MS  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 45 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 07/28/2022  
 Supervised By : Mahesh Dadoda 07/28/2022

Quant Time: Jul 28 05:57:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 25 16:49:17 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 135238              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 237028              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 251657              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 159143              | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964          | 65   | 147002              | 54.081  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 74 - 125 |      | Recovery = 108.160% |         |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 116051              | 51.217  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 102.440% |         |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 443288              | 50.518  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 86 - 113 |      | Recovery = 101.040% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 171967              | 52.806  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 105.620% |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 122717              | 55.300  | ug/l   | 100      |
| 3) Chloromethane             | 1.288          | 50   | 121307              | 55.279  | ug/l   | 98       |
| 4) Vinyl Chloride            | 1.374          | 62   | 114656              | 51.767  | ug/l   | 98       |
| 5) Bromomethane              | 1.599          | 94   | 27087               | 32.899  | ug/l   | 91       |
| 6) Chloroethane              | 1.678          | 64   | 68154               | 53.592  | ug/l   | 94       |
| 7) Trichlorofluoromethane    | 1.880          | 101  | 158831              | 54.430  | ug/l   | 95       |
| 8) Diethyl Ether             | 2.136          | 74   | 60200               | 52.683  | ug/l   | 91       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 112707              | 52.146  | ug/l   | 98       |
| 10) Methyl Iodide            | 2.447          | 142  | 87773               | 41.924  | ug/l   | 96       |
| 11) Tert butyl alcohol       | 2.971          | 59   | 245718              | 370.589 | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 111251              | 52.660  | ug/l   | 94       |
| 13) Acrolein                 | 2.239          | 56   | 39169               | 240.348 | ug/l   | 97       |
| 14) Allyl chloride           | 2.666          | 41   | 225374              | 51.697  | ug/l   | 98       |
| 15) Acrylonitrile            | 3.068          | 53   | 470401              | 287.087 | ug/l   | 99       |
| 16) Acetone                  | 2.386          | 43   | 474792              | 341.173 | ug/l   | 96       |
| 17) Carbon Disulfide         | 2.508          | 76   | 301501              | 52.869  | ug/l   | 99       |
| 18) Methyl Acetate           | 2.709          | 43   | 215063              | 52.247  | ug/l   | 96       |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 438444              | 60.587  | ug/l   | 99       |
| 20) Methylene Chloride       | 2.788          | 84   | 136625              | 57.666  | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 120193              | 51.544  | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.763          | 45   | 478620              | 58.026  | ug/l   | 95       |
| 23) Vinyl Acetate            | 3.727          | 43   | 1944236             | 271.729 | ug/l   | 97       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 239433              | 54.627  | ug/l   | 99       |
| 25) 2-Butanone               | 4.562          | 43   | 700803              | 299.672 | ug/l   | 95       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 100275              | 31.639  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.489          | 96   | 147253              | 53.701  | ug/l   | 98       |
| 28) Bromochloromethane       | 4.903          | 49   | 110089              | 59.454  | ug/l   | 93       |
| 29) Tetrahydrofuran          | 5.007          | 42   | 457859              | 290.986 | ug/l   | 94       |
| 30) Chloroform               | 5.099          | 83   | 243765              | 56.171  | ug/l   | 100      |
| 31) Cyclohexane              | 5.477          | 56   | 216977              | 54.978  | ug/l   | 93       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 195851              | 55.521  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 170668              | 52.339  | ug/l   | 96       |
| 37) Ethyl Acetate            | 4.721          | 43   | 232830              | 50.421  | ug/l   | 97       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 160855              | 52.840  | ug/l   | 95       |
| 39) Methylcyclohexane        | 7.385          | 83   | 213937              | 50.828  | ug/l   | 95       |
| 40) Benzene                  | 6.044          | 78   | 547270              | 53.342  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072722\  
 Data File : VX030261.D  
 Acq On : 28 Jul 2022 03:33  
 Operator : JC/MD  
 Sample : N3884-02MS  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 45 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS

Quant Time: Jul 28 05:57:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 25 16:49:17 2022  
 Response via : Initial Calibration

Manual Integrations  
 APPROVED

Reviewed By :John Carlone 07/28/2022  
 Supervised By :Mahesh Dadoda 07/28/2022

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 139176   | 56.226   | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 193882   | 54.690   | ug/l   | 97       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 369245   | 54.547   | ug/l   | 97       |
| 44) Trichloroethene           | 7.129  | 130  | 133874   | 47.755   | ug/l   | 93       |
| 45) 1,2-Dichloropropane       | 7.434  | 63   | 146420   | 53.509   | ug/l   | 97       |
| 46) Dibromomethane            | 7.586  | 93   | 97582    | 52.439   | ug/l   | 95       |
| 47) Bromodichloromethane      | 7.824  | 83   | 184717   | 54.604   | ug/l   | 98       |
| 48) Methyl methacrylate       | 7.696  | 41   | 190150   | 57.069   | ug/l   | 95       |
| 49) 1,4-Dioxane               | 7.665  | 88   | 80729    | 1049.034 | ug/l   | 96       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 1357306  | 295.347  | ug/l   | 97       |
| 52) Toluene                   | 8.720  | 92   | 331848   | 53.016   | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 185931   | 50.988   | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 8.372  | 75   | 202844   | 51.045   | ug/l   | 96       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 142499   | 53.469   | ug/l   | 96       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 243626   | 57.826   | ug/l   | 93       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 240104   | 54.579   | ug/l   | 100      |
| 59) 2-Hexanone                | 9.433  | 43   | 1079436  | 297.719  | ug/l   | 96       |
| 60) Dibromochloromethane      | 9.525  | 129  | 135682   | 54.997   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 148000   | 53.183   | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 111138   | 41.021   | ug/l   | 95       |
| 65) Chlorobenzene             | 10.079 | 112  | 351200   | 51.375   | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 125379   | 53.953   | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 658214   | 50.293   | ug/l   | 98       |
| 68) m/p-Xylenes               | 10.305 | 106  | 495380   | 106.972  | ug/l   | 94       |
| 69) o-Xylene                  | 10.646 | 106  | 249126   | 54.582   | ug/l   | 95       |
| 70) Styrene                   | 10.659 | 104  | 411350   | 54.856   | ug/l   | 97       |
| 71) Bromoform                 | 10.805 | 173  | 96418    | 56.436   | ug/l # | 98       |
| 73) Isopropylbenzene          | 10.963 | 105  | 637281   | 51.295   | ug/l   | 99       |
| 74) N-amyl acetate            | 10.848 | 43   | 347574   | 57.563   | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 254104   | 55.880   | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 223048m  | 53.171   | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 146409   | 50.209   | ug/l   | 92       |
| 78) n-propylbenzene           | 11.305 | 91   | 785917   | 53.002   | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 458118   | 48.924   | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 542233   | 52.123   | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 11.024 | 75   | 62910    | 44.989   | ug/l   | 93       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 530005   | 51.805   | ug/l   | 96       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 522342   | 54.172   | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 552710   | 53.255   | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 689679   | 52.259   | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 550525   | 52.140   | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 283169   | 49.932   | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 285094   | 49.384   | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 536373   | 54.564   | ug/l   | 99       |
| 90) Hexachloroethane          | 12.542 | 117  | 91529    | 50.348   | ug/l   | 93       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 286107   | 51.400   | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 62965    | 57.732   | ug/l   | 87       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 168052   | 52.061   | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 61366    | 47.345   | ug/l   | 98       |
| 95) Naphthalene               | 13.780 | 128  | 753894   | 59.790   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 174102   | 52.590   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072722\  
Data File : VX030261.D  
Acq On : 28 Jul 2022 03:33  
Operator : JC/MD  
Sample : N3884-02MS  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 45 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS


Manual Integrations  
APPROVED

Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022

Quant Time: Jul 28 05:57:38 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 25 16:49:17 2022  
Response via : Initial Calibration


| Compound                                                             | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------------------------------------------------|------|------|----------|------|-------|----------|
| -----                                                                |      |      |          |      |       |          |
| (#)=qualifier out of range (m)=manual integration (+)=signals summed |      |      |          |      |       |          |





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:168 Resp: 135238  
Ion Ratio Lower Upper  
168 100  
99 56.8 44.0 66.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#2  
Dichlorodifluoromethane  
Concen: 55.300 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 85 Resp: 122717  
Ion Ratio Lower Upper  
85 100  
87 32.8 16.5 49.5









#3  
Chloromethane  
Concen: 55.279 ug/l  
RT: 1.288 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion: 50 Resp: 12130  
Ion Ratio Lower Upper  
50 100  
52 31.9 26.3 39.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#4  
Vinyl Chloride  
Concen: 51.767 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 62 Resp: 114656  
Ion Ratio Lower Upper  
62 100  
64 30.4 25.2 37.8





#5  
Bromomethane  
Concen: 32.899 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 94 Resp: 2708  
Ion Ratio Lower Upper  
94 100  
96 87.3 76.9 115.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#6  
Chloroethane  
Concen: 53.592 ug/l  
RT: 1.678 min Scan# 97  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 64 Resp: 68154  
Ion Ratio Lower Upper  
64 100  
66 30.3 27.2 40.8





#7  
Trichlorofluoromethane  
Concen: 54.430 ug/l  
RT: 1.880 min Scan# 11  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion:101 Resp: 15883  
Ion Ratio Lower Upper  
101 100  
103 65.9 49.7 74.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#8  
Diethyl Ether  
Concen: 52.683 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 74 Resp: 60200  
Ion Ratio Lower Upper  
74 100  
45 111.9 51.5 154.7





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 52.146 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:101 Resp: 11270  
Ion Ratio Lower Upper  
101 100  
85 44.1 34.0 51.0  
151 73.3 59.4 89.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#10  
Methyl Iodide  
Concen: 41.924 ug/l  
RT: 2.447 min Scan# 223  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion:142 Resp: 87773  
Ion Ratio Lower Upper  
142 100  
127 40.6 30.2 45.2  
141 13.9 11.3 16.9









#11  
Tert butyl alcohol  
Concen: 370.589 ug/l  
RT: 2.971 min Scan# 309  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion: 59 Resp: 245718  
Ion Ratio Lower Upper  
59 100  
57 9.1 7.4 11.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#12  
1,1-Dichloroethene  
Concen: 52.660 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 96 Resp: 111251  
Ion Ratio Lower Upper  
96 100  
61 155.9 117.4 176.0  
98 63.0 52.2 78.4





#13  
Acrolein  
Concen: 240.348 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 56 Resp: 39169  
Ion Ratio Lower Upper  
56 100  
55 73.1 56.6 84.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#14  
Allyl chloride  
Concen: 51.697 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 41 Resp: 225374  
Ion Ratio Lower Upper  
41 100  
39 61.5 48.9 73.3  
76 31.1 26.9 40.3





#15  
Acrylonitrile  
Concen: 287.087 ug/l  
RT: 3.068 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 53 Resp: 47040  
Ion Ratio Lower Upper  
53 100  
52 82.2 65.4 98.0  
51 35.1 27.7 41.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#16  
Acetone  
Concen: 341.173 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion: 43 Resp: 474792  
Ion Ratio Lower Upper  
43 100  
58 32.1 27.7 41.5





#17  
Carbon Disulfide  
Concen: 52.869 ug/l  
RT: 2.508 min Scan# 211  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 76 Resp: 30150  
Ion Ratio Lower Upper  
76 100  
78 9.2 7.8 11.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#18  
Methyl Acetate  
Concen: 52.247 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 43 Resp: 215063  
Ion Ratio Lower Upper  
43 100  
74 22.3 19.4 29.2









#19  
Methyl tert-butyl Ether  
Concen: 60.587 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion: 73 Resp: 438444  
Ion Ratio Lower Upper  
73 100  
57 23.4 19.0 28.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#20  
Methylene Chloride  
Concen: 57.666 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 84 Resp: 136625  
Ion Ratio Lower Upper  
84 100  
49 122.3 93.8 140.6  
51 38.8 29.4 44.0  
86 60.9 51.0 76.4





#21  
trans-1,2-Dichloroethene  
Concen: 51.544 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 96 Resp: 12019  
Ion Ratio Lower Upper  
96 100  
61 134.8 105.6 158.4  
98 61.1 50.1 75.1




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#22  
Diisopropyl ether  
Concen: 58.026 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 45 Resp: 478620  
Ion Ratio Lower Upper  
45 100  
43 76.2 65.4 98.0  
87 25.1 22.3 33.5  
59 11.1 9.0 13.4





#23  
 Vinyl Acetate  
 Concen: 271.729 ug/l  
 RT: 3.727 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS




Tgt Ion: 43 Resp: 1944230  
 Ion Ratio Lower Upper  
 43 100  
 86 9.3 8.5 12.7

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 07/28/2022  
 Supervised By :Mahesh Dadoda 07/28/2022




#24  
 1,1-Dichloroethane  
 Concen: 54.627 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. 0.000 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33




Tgt Ion: 63 Resp: 239433  
 Ion Ratio Lower Upper  
 63 100  
 98 6.6 3.5 10.5  
 100 4.7 2.3 6.9





#25  
2-Butanone  
Concen: 299.672 ug/l  
RT: 4.562 min Scan# 571  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 43 Resp: 70080  
Ion Ratio Lower Upper  
43 100  
72 24.5 21.7 32.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#26  
2,2-Dichloropropane  
Concen: 31.639 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 77 Resp: 100275  
Ion Ratio Lower Upper  
77 100  
97 26.4 13.0 39.0





#27  
 cis-1,2-Dichloroethene  
 Concen: 53.701 ug/l  
 RT: 4.489 min Scan# 51  
 Delta R.T. -0.006 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS



Tgt Ion: 96 Resp: 14725  
 Ion Ratio Lower Upper  
 96 100  
 61 138.6 0.0 270.0  
 98 64.0 0.0 127.8




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 07/28/2022  
 Supervised By :Mahesh Dadoda 07/28/2022



#28  
 Bromochloromethane  
 Concen: 59.454 ug/l  
 RT: 4.903 min Scan# 626  
 Delta R.T. 0.000 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33

Tgt Ion: 49 Resp: 110089  
 Ion Ratio Lower Upper  
 49 100  
 129 1.9 0.0 4.4  
 130 72.4 62.8 94.2





#29

Tetrahydrofuran

Concen: 290.986 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 42 Resp: 457859





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 42  | 100   |       |       |
| 72  | 41.5  | 36.5  | 54.7  |
| 71  | 38.7  | 33.9  | 50.9  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#30

Chloroform

Concen: 56.171 ug/l

RT: 5.099 min Scan# 658




Delta R.T. 0.001 min


Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Tgt Ion: 83 Resp: 243765


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 64.6  | 51.5  | 77.3  |





#31  
Cyclohexane  
Concen: 54.978 ug/l  
RT: 5.477 min Scan# 719  
Delta R.T. 0.007 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 56 Resp: 21697  
Ion Ratio Lower Upper  
56 100  
69 28.1 25.4 38.0  
84 79.9 68.9 103.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#32  
1,1,1-Trichloroethane  
Concen: 55.521 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 97 Resp: 195851  
Ion Ratio Lower Upper  
97 100  
99 63.8 51.4 77.0  
61 45.2 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 54.081 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 65 Resp: 147002

Ion Ratio Lower Upper

65 100





67 54.3 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33




Tgt Ion:114 Resp: 237028


Ion Ratio Lower Upper

114 100

63 21.4 0.0 38.4


88 15.1 0.0 31.8





#35  
Dibromofluoromethane  
Concen: 51.217 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:113 Resp: 11605  
Ion Ratio Lower Upper  
113 100  
111 103.1 83.2 124.8  
192 17.8 14.8 22.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#36  
1,1-Dichloropropene  
Concen: 52.339 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 75 Resp: 170668  
Ion Ratio Lower Upper  
75 100  
110 35.3 19.3 57.9  
77 31.1 24.4 36.6





#37

Ethyl Acetate

Concen: 50.421 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.001 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 43 Resp: 232830





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 13.3  | 11.7  | 17.5  |
| 70  | 10.2  | 9.3   | 13.9  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#38

Carbon Tetrachloride

Concen: 52.840 ug/l

RT: 5.684 min Scan# 754




Delta R.T. 0.001 min


Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Tgt Ion:117 Resp: 160855


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 98.8  | 74.1  | 111.1 |
| 121 | 31.6  | 25.8  | 38.8  |





#39  
Methylcyclohexane  
Concen: 50.828 ug/l  
RT: 7.385 min Scan# 1033  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 83 Resp: 21393  
Ion Ratio Lower Upper  
83 100  
55 78.5 57.7 86.5  
98 44.5 36.5 54.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#40  
Benzene  
Concen: 53.342 ug/l  
RT: 6.044 min Scan# 813  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 78 Resp: 547270  
Ion Ratio Lower Upper  
78 100  
77 23.7 18.0 27.0





#41  
Methacrylonitrile  
Concen: 56.226 ug/l  
RT: 4.928 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 41 Resp: 139170  
Ion Ratio Lower Upper  
41 100  
39 55.5 42.9 64.3  
67 68.7 60.1 90.1  
52 27.8 22.6 34.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#42  
1,2-Dichloroethane  
Concen: 54.690 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion: 62 Resp: 193882  
Ion Ratio Lower Upper  
62 100  
98 8.9 0.0 20.2





#43  
Isopropyl Acetate  
Concen: 54.547 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 43 Resp: 369249  
Ion Ratio Lower Upper  
43 100  
61 19.2 16.5 24.7  
87 12.5 11.1 16.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#44  
Trichloroethene  
Concen: 47.755 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Tgt Ion:130 Resp: 133874  
Ion Ratio Lower Upper  
130 100  
95 104.1 0.0 194.8





#45

1,2-Dichloropropane

Concen: 53.509 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.001 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 63 Resp: 146420

Ion Ratio Lower Upper

63 100





65 29.6 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#46

Dibromomethane

Concen: 52.439 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33




Tgt Ion: 93 Resp: 97582


Ion Ratio Lower Upper

93 100

95 82.8 67.4 101.0


174 87.3 75.6 113.4





#47  
Bromodichloromethane  
Concen: 54.604 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.005 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 83 Resp: 18471  
Ion Ratio Lower Upper  
83 100  
85 65.6 53.5 80.3  
127 7.9 6.6 10.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#48  
Methyl methacrylate  
Concen: 57.069 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion: 41 Resp: 190150  
Ion Ratio Lower Upper  
41 100  
69 79.4 68.8 103.2  
39 54.8 44.6 66.8





#49  
1,4-Dioxane  
Concen: 1049.034 ug/l  
RT: 7.665 min Scan# 1079  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 88 Resp: 80729  
Ion Ratio Lower Upper  
88 100  
43 32.1 23.8 35.8  
58 70.8 53.8 80.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#50  
Toluene-d8  
Concen: 50.518 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion: 98 Resp: 443288  
Ion Ratio Lower Upper  
98 100  
100 64.1 52.3 78.5





#51  
4-Methyl-2-Pentanone  
Concen: 295.347 ug/l  
RT: 8.580 min Scan# 11  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 43 Resp: 1357300  
Ion Ratio Lower Upper  
43 100  
58 37.4 31.4 47.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#52  
Toluene  
Concen: 53.016 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 92 Resp: 331848  
Ion Ratio Lower Upper  
92 100  
91 176.2 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 50.988 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 75 Resp: 18593

Ion Ratio Lower Upper

75 100





77 31.2 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#54

cis-1,3-Dichloropropene

Concen: 51.045 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33




Tgt Ion: 75 Resp: 202844


Ion Ratio Lower Upper

75 100

77 31.4 25.0 37.4

39 51.3 37.5 56.3





#55

1,1,2-Trichloroethane

Concen: 53.469 ug/l

RT: 9.153 min Scan# 1324

Delta R.T. -0.005 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 97 Resp: 142499

Ion Ratio Lower Upper

97 100

83 90.7 68.9 103.3

85 56.5 44.0 66.0





99 63.6 48.3 72.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#56

Ethyl methacrylate

Concen: 57.826 ug/l

RT: 9.122 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33




Tgt Ion: 69 Resp: 243626


Ion Ratio Lower Upper

69 100

41 71.3 52.3 78.5

39 38.3 28.2 42.4





#57

1,3-Dichloropropane

Concen: 54.579 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion: 76 Resp: 240104

Ion Ratio Lower Upper

76 100





78 32.2 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#59

2-Hexanone

Concen: 297.719 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX030261.D




Acq: 28 Jul 2022 03:33


Tgt Ion: 43 Resp: 1079436

Ion Ratio Lower Upper

43 100


58 52.3 27.6 82.8





#60  
Dibromochloromethane  
Concen: 54.997 ug/l  
RT: 9.525 min Scan# 11  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:129 Resp: 13568  
Ion Ratio Lower Upper  
129 100  
127 77.5 38.5 115.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#61  
1,2-Dibromoethane  
Concen: 53.183 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:107 Resp: 148000  
Ion Ratio Lower Upper  
107 100  
109 95.2 76.6 114.8





#62  
4-Bromofluorobenzene  
Concen: 52.806 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 95 Resp: 17196  
Ion Ratio Lower Upper  
95 100  
174 69.0 0.0 149.0  
176 67.3 0.0 146.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:117 Resp: 251657  
Ion Ratio Lower Upper  
117 100  
82 59.1 42.7 64.1  
119 31.7 25.4 38.2





#64  
Tetrachloroethene  
Concen: 41.021 ug/l  
RT: 9.275 min Scan# 11113  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion:164 Resp: 11113  
Ion Ratio Lower Upper  
164 100  
166 127.8 98.2 147.4  
129 102.9 80.8 121.2  
131 101.8 75.5 113.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#65  
Chlorobenzene  
Concen: 51.375 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion:112 Resp: 351200  
Ion Ratio Lower Upper  
112 100  
114 32.0 24.6 36.8





#66  
1,1,1,2-Tetrachloroethane  
Concen: 53.953 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:131 Resp: 125379  
Ion Ratio Lower Upper  
131 100  
133 96.1 46.9 140.8  
119 69.0 34.4 103.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#67  
Ethyl Benzene  
Concen: 50.293 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion: 91 Resp: 658214  
Ion Ratio Lower Upper  
91 100  
106 30.6 25.5 38.3





#68  
m/p-Xylenes  
Concen: 106.972 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion:106 Resp: 495380  
Ion Ratio Lower Upper  
106 100  
91 206.4 158.2 237.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#69  
o-Xylene  
Concen: 54.582 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Tgt Ion:106 Resp: 249126  
Ion Ratio Lower Upper  
106 100  
91 215.2 103.8 311.6





#70

Styrene

Concen: 54.856 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.001 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS





Tgt Ion:104 Resp: 411350  
 Ion Ratio Lower Upper  
 104 100  
 78 51.6 37.8 56.8  
 103 54.9 44.5 66.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#71

Bromoform

Concen: 56.436 ug/l




RT: 10.805 min Scan# 1594


Delta R.T. -0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33


Tgt Ion:173 Resp: 96418  
 Ion Ratio Lower Upper  
 173 100  
 175 49.2 23.8 71.2  
 254 0.0 0.1 0.1#





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:152 Resp: 15914  
Ion Ratio Lower Upper  
152 100  
115 86.5 40.5 121.5  
150 174.9 0.0 347.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#73  
Isopropylbenzene  
Concen: 51.295 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:105 Resp: 637281  
Ion Ratio Lower Upper  
105 100  
120 26.4 13.5 40.4





#74  
N-amy1 acetate  
Concen: 57.563 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion: 43 Resp: 347574  
Ion Ratio Lower Upper  
43 100  
70 38.5 32.3 48.5  
55 22.9 18.4 27.6  
61 22.9 19.2 28.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#75  
1,1,2,2-Tetrachloroethane  
Concen: 55.880 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 83 Resp: 254104  
Ion Ratio Lower Upper  
83 100  
131 9.0 4.5 13.7  
85 64.8 32.5 97.5









#76  
1,2,3-Trichloropropane  
Concen: 53.171 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion: 75 Resp: 223048  
Ion Ratio Lower Upper  
75 100  
77 38.3 19.8 59.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#77  
Bromobenzene  
Concen: 50.209 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion:156 Resp: 146409  
Ion Ratio Lower Upper  
156 100  
77 181.8 82.7 247.9  
158 96.7 47.5 142.5





#78  
n-propylbenzene  
Concen: 53.002 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.005 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS



Tgt Ion: 91 Resp: 785917  
Ion Ratio Lower Upper  
91 100  
120 22.3 11.6 34.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#79  
2-Chlorotoluene  
Concen: 48.924 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. 0.001 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Tgt Ion: 91 Resp: 458118  
Ion Ratio Lower Upper  
91 100  
126 33.5 17.4 52.3





#80

1,3,5-Trimethylbenzene

Concen: 52.123 ug/l

RT: 11.451 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33

Instrument :

MSVOA\_X

ClientSampleId :

BEM-MW-04MS

Tgt Ion:105 Resp: 54223

Ion Ratio Lower Upper

105 100





120 47.3 24.4 73.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 07/28/2022

Supervised By :Mahesh Dadoda 07/28/2022



#81

trans-1,4-Dichloro-2-butene

Concen: 44.989 ug/l

RT: 11.024 min Scan# 1630

Delta R.T. 0.000 min

Lab File: VX030261.D

Acq: 28 Jul 2022 03:33




Tgt Ion: 75 Resp: 62910


Ion Ratio Lower Upper

75 100

53 101.3 74.2 111.4

89 45.4 34.8 52.2









#82  
4-Chlorotoluene  
Concen: 51.805 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion: 91 Resp: 53000  
Ion Ratio Lower Upper  
91 100  
126 28.7 15.3 45.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#83  
tert-Butylbenzene  
Concen: 54.172 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion:119 Resp: 522342  
Ion Ratio Lower Upper  
119 100  
91 60.9 28.3 84.9  
134 22.1 11.2 33.5





#84  
1,2,4-Trimethylbenzene  
Concen: 53.255 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:105 Resp: 552710  
Ion Ratio Lower Upper  
105 100  
120 44.5 22.9 68.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#85  
sec-Butylbenzene  
Concen: 52.259 ug/l  
RT: 11.890 min Scan# 1772  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:105 Resp: 689679  
Ion Ratio Lower Upper  
105 100  
134 19.4 10.0 30.0





#86  
p-Isopropyltoluene  
Concen: 52.140 ug/l  
RT: 12.012 min Scan# 119  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:119 Resp: 55052  
Ion Ratio Lower Upper  
119 100  
134 26.3 13.9 41.6  
91 25.2 11.8 35.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#87  
1,3-Dichlorobenzene  
Concen: 49.932 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:146 Resp: 283169  
Ion Ratio Lower Upper  
146 100  
111 42.9 20.3 60.8  
148 62.9 31.8 95.3





#88  
1,4-Dichlorobenzene  
Concen: 49.384 ug/l  
RT: 12.042 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:146 Resp: 285094  
Ion Ratio Lower Upper  
146 100  
111 41.9 20.3 60.9  
148 63.7 32.3 96.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#89  
n-Butylbenzene  
Concen: 54.564 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33



Tgt Ion: 91 Resp: 536373  
Ion Ratio Lower Upper  
91 100  
92 52.2 26.2 78.5  
134 24.4 12.6 37.8









#90  
Hexachloroethane  
Concen: 50.348 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33

Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS

Tgt Ion:117 Resp: 91529  
Ion Ratio Lower Upper  
117 100  
201 63.0 34.2 102.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022



#91  
1,2-Dichlorobenzene  
Concen: 51.400 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. -0.006 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33


Tgt Ion:146 Resp: 286107  
Ion Ratio Lower Upper  
146 100  
111 44.7 20.7 62.1  
148 63.7 32.0 96.2





#92  
 1,2-Dibromo-3-Chloropropane  
 Concen: 57.732 ug/l  
 RT: 12.945 min Scan# 1945  
 Delta R.T. 0.001 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS




Tgt Ion: 75 Resp: 6296  
 Ion Ratio Lower Upper  
 75 100  
 155 77.5 45.0 134.8  
 157 99.7 56.9 170.7

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 07/28/2022  
 Supervised By :Mahesh Dadoda 07/28/2022




#93  
 1,2,4-Trichlorobenzene  
 Concen: 52.061 ug/l  
 RT: 13.591 min Scan# 2051  
 Delta R.T. -0.000 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33




Tgt Ion:180 Resp: 168052  
 Ion Ratio Lower Upper  
 180 100  
 182 95.0 48.2 144.6  
 145 33.7 15.9 47.6





#94  
Hexachlorobutadiene  
Concen: 47.345 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Instrument :  
MSVOA\_X  
ClientSampleId :  
BEM-MW-04MS




Tgt Ion:225 Resp: 61360  
Ion Ratio Lower Upper  
225 100  
223 62.8 30.6 91.6  
227 63.3 31.9 95.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 07/28/2022  
Supervised By :Mahesh Dadoda 07/28/2022




#95  
Naphthalene  
Concen: 59.790 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX030261.D  
Acq: 28 Jul 2022 03:33




Tgt Ion:128 Resp: 753894  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.2 15.2  
129 11.0 8.6 13.0





#96  
 1,2,3-Trichlorobenzene  
 Concen: 52.590 ug/l  
 RT: 13.963 min Scan# 2112  
 Delta R.T. 0.001 min  
 Lab File: VX030261.D  
 Acq: 28 Jul 2022 03:33

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BEM-MW-04MS



Tgt Ion:180 Resp: 174102  
 Ion Ratio Lower Upper  
 180 100  
 182 94.9 47.0 141.0  
 145 35.6 17.0 50.9

Manual Integrations  
 APPROVED

Reviewed By :John Carlone 07/28/2022  
 Supervised By :Mahesh Dadoda 07/28/2022

