

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036668.D
 Acq On : 27 Jul 2023 11:45
 Operator : JC/MD
 Sample : VX0727MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0727MBL01

Quant Time: Jul 28 06:08:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	153815	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	274736	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105043	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102089	44.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.720%
35) Dibromofluoromethane	5.385	113	87417	46.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	8.647	98	326600	48.771	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.079	95	119380	47.251	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036668.D
Acq On : 27 Jul 2023 11:45
Operator : JC/MD
Sample : VX0727MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0727MBL01

Quant Time: Jul 28 06:08:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

