

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036676.D
 Acq On : 27 Jul 2023 15:04
 Operator : JC/MD
 Sample : 03767-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MMC-COMPRE

Quant Time: Jul 28 06:12:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92471	44.533	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.060%
35) Dibromofluoromethane	5.391	113	79053	46.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	8.647	98	300560	49.631	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.260%
62) 4-Bromofluorobenzene	11.079	95	110468	48.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.700%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	25064	29.088	ug/l	90
20) Methylene Chloride	2.788	84	4058	2.118	ug/l	89
25) 2-Butanone	4.580	43	2884	2.063	ug/l	96
29) Tetrahydrofuran	5.019	42	1946	2.159	ug/l #	77
30) Chloroform	5.099	83	25786	7.651	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036676.D
Acq On : 27 Jul 2023 15:04
Operator : JC/MD
Sample : 03767-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MMC-COMPRE

Quant Time: Jul 28 06:12:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

