

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036678.D
 Acq On : 27 Jul 2023 15:53
 Operator : JC/MD
 Sample : 03782-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-6I

Quant Time: Jul 28 06:13:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92658	45.298	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.600%
35) Dibromofluoromethane	5.391	113	78829	45.900	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.800%
50) Toluene-d8	8.653	98	299655	49.201	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.400%
62) 4-Bromofluorobenzene	11.079	95	109309	47.571	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4610	5.431	ug/l	90
20) Methylene Chloride	2.782	84	3733	1.978	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036678.D
Acq On : 27 Jul 2023 15:53
Operator : JC/MD
Sample : 03782-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-6I

Quant Time: Jul 28 06:13:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

