

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036683.D
 Acq On : 27 Jul 2023 17:54
 Operator : JC/MD
 Sample : 03782-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01

Quant Time: Jul 28 06:15:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81660	45.343	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.680%
35) Dibromofluoromethane	5.391	113	69941	46.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.640%
50) Toluene-d8	8.647	98	257380	48.585	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	93438	46.750	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18345	24.548	ug/l #	86
20) Methylene Chloride	2.788	84	2303	1.386	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036683.D
Acq On : 27 Jul 2023 17:54
Operator : JC/MD
Sample : 03782-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01

Quant Time: Jul 28 06:15:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

