

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036688.D
 Acq On : 27 Jul 2023 19:55
 Operator : JC/MD
 Sample : 03811-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-2I

Quant Time: Jul 28 06:16:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87148	45.617	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.240%
35) Dibromofluoromethane	5.391	113	73622	45.725	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.647	98	282149	49.415	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.840%
62) 4-Bromofluorobenzene	11.079	95	105770	49.099	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.200%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3664	4.622	ug/l	94
44) Trichloroethene	7.129	130	2193	1.202	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036688.D
Acq On : 27 Jul 2023 19:55
Operator : JC/MD
Sample : 03811-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-2I

Quant Time: Jul 28 06:16:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

