

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036691.D
 Acq On : 27 Jul 2023 21:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 06:17:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.661	0.580	12.3	70	0.00
3 P	Chloromethane	0.601	0.553	8.0	74	0.02
4 C	Vinyl Chloride	0.612	0.585	4.4#	74	0.00
5 T	Bromomethane	0.365	0.401	-9.9	87	-0.02
6 T	Chloroethane	0.297	0.274	7.7	67	-0.02
7 T	Trichlorofluoromethane	0.945	0.963	-1.9	79	0.00
8 T	Diethyl Ether	0.365	0.358	1.9	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.547	3.5	74	0.00
10 T	Methyl Iodide	0.589	0.621	-5.4	76	0.00
11 T	Tert butyl alcohol	0.175	0.211	-20.6	93	-0.02
12 CM	1,1-Dichloroethene	0.560	0.539	3.8#	75	0.00
13 T	Acrolein	0.100	0.106	-6.0	84	0.00
14 T	Allyl chloride	1.102	0.958	13.1	68	0.00
15 T	Acrylonitrile	0.352	0.386	-9.7	83	0.00
16 T	Acetone	0.307	0.332	-8.1	85	0.00
17 T	Carbon Disulfide	1.545	1.331	13.9	68	0.00
18 T	Methyl Acetate	1.291	1.379	-6.8	81	0.00
19 T	Methyl tert-butyl Ether	2.104	2.155	-2.4	78	0.00
20 T	Methylene Chloride	0.683	0.640	6.3	76	0.00
21 T	trans-1,2-Dichloroethene	0.626	0.600	4.2	75	0.00
22 T	Diisopropyl ether	2.093	2.011	3.9	74	0.00
23 T	Vinyl Acetate	1.606	1.587	1.2	74	0.00
24 P	1,1-Dichloroethane	1.174	1.162	1.0	76	0.00
25 T	2-Butanone	0.498	0.549	-10.2	84	0.00
26 T	2,2-Dichloropropane	1.106	0.725	34.4#	51	0.00
27 T	cis-1,2-Dichloroethene	0.743	0.727	2.2	77	0.00
28 T	Bromochloromethane	0.568	0.548	3.5	73	0.00
29 T	Tetrahydrofuran	0.321	0.345	-7.5	81	0.00
30 C	Chloroform	1.201	1.214	-1.1#	78	0.00
31 T	Cyclohexane	1.024	0.921	10.1	70	0.00
32 T	1,1,1-Trichloroethane	1.100	1.083	1.5	76	0.00
33 S	1,2-Dichloroethane-d4	0.740	0.717	3.1	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.344	0.338	1.7	78	0.00
36 T	1,1-Dichloropropene	0.512	0.474	7.4	73	0.00
37 T	Ethyl Acetate	0.580	0.596	-2.8	79	0.00
38 T	Carbon Tetrachloride	0.545	0.533	2.2	76	0.00
39 T	Methylcyclohexane	0.637	0.575	9.7	71	0.00
40 TM	Benzene	1.490	1.444	3.1	75	0.00
41 T	Methacrylonitrile	0.310	0.320	-3.2	78	0.00
42 TM	1,2-Dichloroethane	0.545	0.550	-0.9	77	0.00
43 T	Isopropyl Acetate	0.998	0.975	2.3	76	0.00
44 TM	Trichloroethene	0.390	0.382	2.1	76	0.00
45 C	1,2-Dichloropropane	0.398	0.385	3.3#	75	0.00
46 T	Dibromomethane	0.278	0.284	-2.2	79	0.00
47 T	Bromodichloromethane	0.572	0.567	0.9	77	0.00
48 T	Methyl methacrylate	0.464	0.466	-0.4	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
 Data File : VX036691.D
 Acq On : 27 Jul 2023 21:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 06:17:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.012	-33.3#	97	-0.04
50 S	Toluene-d8	1.219	1.192	2.2	76	0.00
51 T	4-Methyl-2-Pentanone	0.552	0.595	-7.8	81	0.00
52 CM	Toluene	0.934	0.922	1.3#	76	0.00
53 T	t-1,3-Dichloropropene	0.646	0.595	7.9	70	0.00
54 T	cis-1,3-Dichloropropene	0.682	0.631	7.5	71	0.00
55 T	1,1,2-Trichloroethane	0.393	0.394	-0.3	78	0.00
56 T	Ethyl methacrylate	0.639	0.659	-3.1	77	0.00
57 T	1,3-Dichloropropane	0.654	0.656	-0.3	76	0.00
58 T	2-Chloroethyl Vinyl ether	0.315	0.316	-0.3	77	0.00
59 T	2-Hexanone	0.410	0.458	-11.7	84	0.00
60 T	Dibromochloromethane	0.431	0.442	-2.6	77	0.00
61 T	1,2-Dibromoethane	0.407	0.416	-2.2	78	0.00
62 S	4-Bromofluorobenzene	0.460	0.456	0.9	74	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.373	0.360	3.5	75	0.00
65 PM	Chlorobenzene	1.102	1.084	1.6	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.435	-2.1	79	0.00
67 C	Ethyl Benzene	2.013	1.969	2.2#	76	0.00
68 T	m/p-Xylenes	0.751	0.747	0.5	76	0.00
69 T	o-Xylene	0.750	0.750	0.0	76	0.00
70 T	Styrene	1.237	1.249	-1.0	76	0.00
71 P	Bromoform	0.345	0.358	-3.8	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	4.138	3.890	6.0	77	0.00
74 T	N-amyl acetate	1.930	1.824	5.5	76	0.00
75 P	1,1,2,2-Tetrachloroethane	1.398	1.379	1.4	82	0.00
76 T	1,2,3-Trichloropropane	1.176	1.151	2.1	79	0.00
77 T	Bromobenzene	0.942	0.915	2.9	77	0.00
78 T	n-propylbenzene	4.786	4.521	5.5	75	0.00
79 T	2-Chlorotoluene	2.900	2.718	6.3	76	0.00
80 T	1,3,5-Trimethylbenzene	3.462	3.292	4.9	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.519	0.447	13.9	68	0.00
82 T	4-Chlorotoluene	3.304	3.136	5.1	77	0.00
83 T	tert-Butylbenzene	3.401	3.333	2.0	79	0.00
84 T	1,2,4-Trimethylbenzene	3.462	3.275	5.4	76	0.00
85 T	sec-Butylbenzene	4.241	4.082	3.7	77	0.00
86 T	p-Isopropyltoluene	3.490	3.400	2.6	77	0.00
87 T	1,3-Dichlorobenzene	1.745	1.713	1.8	78	0.00
88 T	1,4-Dichlorobenzene	1.732	1.717	0.9	79	0.00
89 T	n-Butylbenzene	3.124	3.012	3.6	76	0.00
90 T	Hexachloroethane	0.707	0.683	3.4	77	0.00
91 T	1,2-Dichlorobenzene	1.684	1.717	-2.0	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.325	0.336	-3.4	86	0.00
93 T	1,2,4-Trichlorobenzene	1.038	1.061	-2.2	81	0.00
94 T	Hexachlorobutadiene	0.429	0.404	5.8	77	0.00
95 T	Naphthalene	3.616	3.853	-6.6	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072723\
Data File : VX036691.D
Acq On : 27 Jul 2023 21:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 28 06:17:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.023	1.067	-4.3	83	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6