

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072724\
 Data File : VX042566.D
 Acq On : 25 Jul 2024 15:03
 Operator : JC/MD
 Sample : VX0727MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK697

Quant Time: Jul 26 00:19:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 26 00:18:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	217344	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	197223	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	95280	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	42499	33.594	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.180%
7) Chloroethane-d5	1.660	69	33926	42.592	ug/L	0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.180%
11) 1,1-Dichloroethene-d2	2.300	65	23160	41.830	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.660%
21) 2-Butanone-d5	4.452	46	86753	95.159	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.160%
24) Chloroform-d	5.056	84	110065	43.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.540%
26) 1,2-Dichloroethane-d4	5.952	65	70808	47.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.320%
32) Benzene-d6	5.970	84	223811	43.904	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	7.306	67	68011	44.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.660%
41) Toluene-d8	8.647	98	205121	43.178	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.360%
43) trans-1,3-Dichloroprop...	8.952	79	30506	46.009	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.020%
47) 2-Hexanone-d5	9.384	63	64702	92.913	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.910%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	99206	43.741	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.480%
66) 1,2-Dichlorobenzene-d4	12.317	152	85986	46.463	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072724\
Data File : VX042566.D
Acq On : 25 Jul 2024 15:03
Operator : JC/MD
Sample : VX0727MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK697

Quant Time: Jul 26 00:19:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 26 00:18:59 2024
Response via : Initial Calibration

