

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072724\
 Data File : VX042588.D
 Acq On : 25 Jul 2024 23:32
 Operator : JC/MD
 Sample : P3328-08ME
 Misc : 9.62g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E1ZW3ME

Quant Time: Jul 26 01:51:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 26 00:18:59 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 07/26/2024
 Supervised By : Mahesh Dadoda 07/26/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	235129	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	212386	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	101338	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	28026	20.478	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	40.960%#		
7) Chloroethane-d5	1.648	69	17708m	20.550	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	41.100%#		
11) 1,1-Dichloroethene-d2	2.276	65	12145	20.276	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	40.560%#		
21) 2-Butanone-d5	4.489	46	46137	46.779	ug/L	0.03
Spiked Amount 100.000	Range 40 - 130		Recovery =	46.780%		
24) Chloroform-d	5.056	84	68166	24.772	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	49.540%#		
26) 1,2-Dichloroethane-d4	5.952	65	44948	27.674	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	55.340%#		
32) Benzene-d6	5.964	84	141413	25.760	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	51.520%#		
36) 1,2-Dichloropropane-d6	7.305	67	42672	25.829	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	51.660%#		
41) Toluene-d8	8.647	98	128132	25.046	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	50.100%#		
43) trans-1,3-Dichloroprop...	8.951	79	15956	22.347	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	44.700%#		
47) 2-Hexanone-d5	9.390	63	35216	46.960	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	46.960%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	60278	24.680	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	49.360%#		
66) 1,2-Dichlorobenzene-d4	12.323	152	54280	27.577	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	55.160%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072724\
Data File : VX042588.D
Acq On : 25 Jul 2024 23:32
Operator : JC/MD
Sample : P3328-08ME
Misc : 9.62g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E1ZW3ME

Quant Time: Jul 26 01:51:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 26 00:18:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 07/26/2024
Supervised By :Mahesh Dadoda 07/26/2024

