

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072820\
 Data File : VX017650.D
 Acq On : 28 Jul 2020 14:00
 Operator : JC/SP
 Sample : L3371-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-200720

Quant Time: Jul 29 03:20:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	493037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	823053	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	868156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	460264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	344899	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	5.47	113	280909	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	8.70	98	1010360	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	11.12	95	445644	56.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.16%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	41391	6.985	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072820\
Data File : VX017650.D
Acq On : 28 Jul 2020 14:00
Operator : JC/SP
Sample : L3371-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-200720

Quant Time: Jul 29 03:20:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

