

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX072820\
 Data File : VX017653.D
 Acq On : 28 Jul 2020 15:09
 Operator : JC/SP
 Sample : L3413-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40956DL

Quant Time: Jul 29 04:02:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	532793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	883426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	927293	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	531337	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	359941	52.59	ug/l	0.00
Spiked Amount			Recovery	=	105.18%	
35) Dibromofluoromethane	5.47	113	299880	50.75	ug/l	0.00
Spiked Amount			Recovery	=	101.50%	
50) Toluene-d8	8.70	98	1100846	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	11.12	95	504206	59.11	ug/l	0.00
Spiked Amount			Recovery	=	118.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	9583	6.644	ug/l	95
16) Acetone	2.42	43	602182	199.910	ug/l	95
18) Methyl Acetate	2.75	43	10065	1.419	ug/l #	75
19) Methyl tert-butyl Ether	3.17	73	53613	2.815	ug/l	98
22) Diisopropyl ether	3.82	45	14259	0.776	ug/l #	87
25) 2-Butanone	4.64	43	144776	33.832	ug/l	100
37) Ethyl Acetate	4.79	43	1354726	146.132	ug/l	98
40) Benzene	6.12	78	1034387	42.583	ug/l	99
52) Toluene	8.77	92	1088970	69.902	ug/l	99
67) Ethyl Benzene	10.24	91	126281	3.769	ug/l	97
68) m/p-Xylenes	10.34	106	183159	14.111	ug/l	94
69) o-Xylene	10.68	106	119427	9.567	ug/l	97
78) n-propylbenzene	11.35	91	20560	0.500	ug/l	88
80) 1,3,5-Trimethylbenzene	11.49	105	46130	1.482	ug/l	100
84) 1,2,4-Trimethylbenzene	11.79	105	244998	7.915	ug/l	99
95) Naphthalene	13.82	128	157495	4.488	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072820\
Data File : VX017653.D
Acq On : 28 Jul 2020 15:09
Operator : JC/SP
Sample : L3413-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40956DL

Quant Time: Jul 29 04:02:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

