

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072820\
 Data File : VX017661.D
 Acq On : 28 Jul 2020 18:21
 Operator : JC/SP
 Sample : L3457-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SHB-MOTOR-ROOM

Quant Time: Jul 29 03:52:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	481190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	813818	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	856659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	471125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	337226	54.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.12%	
35) Dibromofluoromethane	5.47	113	280076	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
50) Toluene-d8	8.70	98	1016351	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	11.12	95	448072	57.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.04%	
Target Compounds						
16) Acetone	2.42	43	15220	5.595	ug/l #	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072820\
Data File : VX017661.D
Acq On : 28 Jul 2020 18:21
Operator : JC/SP
Sample : L3457-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SHB-MOTOR-ROOM

Quant Time: Jul 29 03:52:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

