

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072821\
 Data File : VX023489.D
 Acq On : 28 Jul 2021 13:35
 Operator : JC/MD
 Sample : M3184-02
 Misc : 14.35g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20210482-68

Quant Time: Jul 28 19:44:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	308159	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110886	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	120593	52.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.420%
35) Dibromofluoromethane	5.397	113	93522	48.539	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.080%
50) Toluene-d8	8.653	98	373869	52.814	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.620%
62) 4-Bromofluorobenzene	11.085	95	127066	52.788	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072821\
Data File : VX023489.D
Acq On : 28 Jul 2021 13:35
Operator : JC/MD
Sample : M3184-02
Misc : 14.35g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20210482-68

Quant Time: Jul 28 19:44:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

