

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072821\
 Data File : VX023490.D
 Acq On : 28 Jul 2021 14:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 19:44:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	52.068	-4.1	93	0.00
3 P	Chloromethane	50.000	50.668	-1.3	93	0.00
4 C	Vinyl Chloride	50.000	49.748	0.5#	90	0.00
5 T	Bromomethane	50.000	46.648	6.7	92	0.00
6 T	Chloroethane	50.000	51.794	-3.6	93	0.00
7 T	Trichlorofluoromethane	50.000	51.265	-2.5	93	0.00
8 T	Diethyl Ether	50.000	49.534	0.9	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.419	-2.8	94	0.00
10 T	Methyl Iodide	50.000	53.833	-7.7	93	0.00
11 T	Tert butyl alcohol	250.000	248.202	0.7	91	0.00
12 CM	1,1-Dichloroethene	50.000	49.104	1.8#	90	0.00
13 T	Acrolein	250.000	231.980	7.2	88	0.00
14 T	Allyl chloride	50.000	49.625	0.8	91	0.00
15 T	Acrylonitrile	250.000	246.755	1.3	88	0.00
16 T	Acetone	250.000	240.906	3.6	93	0.00
17 T	Carbon Disulfide	50.000	43.740	12.5	80	0.00
18 T	Methyl Acetate	50.000	50.362	-0.7	92	0.00
19 T	Methyl tert-butyl Ether	50.000	51.427	-2.9	92	0.00
20 T	Methylene Chloride	50.000	48.145	3.7	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.222	1.6	90	0.00
22 T	Diisopropyl ether	50.000	51.584	-3.2	92	0.00
23 T	Vinyl Acetate	250.000	257.245	-2.9	90	0.00
24 P	1,1-Dichloroethane	50.000	50.933	-1.9	92	0.00
25 T	2-Butanone	250.000	245.887	1.6	88	0.00
26 T	2,2-Dichloropropane	50.000	51.892	-3.8	93	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.385	-0.8	91	0.00
28 T	Bromochloromethane	50.000	48.143	3.7	90	0.00
29 T	Tetrahydrofuran	250.000	244.318	2.3	87	0.00
30 C	Chloroform	50.000	50.505	-1.0#	92	0.00
31 T	Cyclohexane	50.000	48.053	3.9	90	0.00
32 T	1,1,1-Trichloroethane	50.000	51.699	-3.4	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.100	3.8	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	48.946	2.1	92	0.00
36 T	1,1-Dichloropropene	50.000	52.333	-4.7	92	0.00
37 T	Ethyl Acetate	50.000	49.955	0.1	90	0.00
38 T	Carbon Tetrachloride	50.000	51.905	-3.8	91	0.00
39 T	Methylcyclohexane	50.000	52.504	-5.0	92	0.00
40 TM	Benzene	50.000	51.466	-2.9	91	0.00
41 T	Methacrylonitrile	50.000	51.398	-2.8	92	0.00
42 TM	1,2-Dichloroethane	50.000	53.177	-6.4	94	0.00
43 T	Isopropyl Acetate	50.000	51.593	-3.2	90	0.00
44 TM	Trichloroethene	50.000	51.779	-3.6	91	0.00
45 C	1,2-Dichloropropane	50.000	52.148	-4.3#	91	0.00
46 T	Dibromomethane	50.000	52.031	-4.1	90	0.00
47 T	Bromodichloromethane	50.000	52.857	-5.7	90	0.00
48 T	Methyl methacrylate	50.000	54.047	-8.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072821\
 Data File : VX023490.D
 Acq On : 28 Jul 2021 14:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 19:44:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1095.603	-9.6	97	0.00
50 S	Toluene-d8	50.000	50.210	-0.4	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	275.224	-10.1	93	0.00
52 CM	Toluene	50.000	53.968	-7.9#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	48.934	2.1	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.696	-7.4	90	0.00
55 T	1,1,2-Trichloroethane	50.000	55.111	-10.2	93	0.00
56 T	Ethyl methacrylate	50.000	55.843	-11.7	93	0.00
57 T	1,3-Dichloropropane	50.000	53.649	-7.3	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.403	-3.0	91	0.00
59 T	2-Hexanone	250.000	287.330	-14.9	97	0.00
60 T	Dibromochloromethane	50.000	48.969	2.1	90	0.00
61 T	1,2-Dibromoethane	50.000	55.183	-10.4	93	0.00
62 S	4-Bromofluorobenzene	50.000	57.132	-14.3	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	51.103	-2.2	95	0.00
65 PM	Chlorobenzene	50.000	51.994	-4.0	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.228	1.5	93	0.00
67 C	Ethyl Benzene	50.000	53.544	-7.1#	97	0.00
68 T	m/p-Xylenes	100.000	107.879	-7.9	98	0.00
69 T	o-Xylene	50.000	54.272	-8.5	100	0.00
70 T	Styrene	50.000	56.637	-13.3	101	0.00
71 P	Bromoform	50.000	46.689	6.6	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	51.423	-2.8	100	0.00
74 T	N-ethyl acetate	50.000	52.049	-4.1	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.314	-0.6	102	0.00
76 T	1,2,3-Trichloropropane	50.000	53.037	-6.1	104	0.00
77 T	Bromobenzene	50.000	52.902	-5.8	103	0.00
78 T	n-propylbenzene	50.000	52.984	-6.0	102	0.00
79 T	2-Chlorotoluene	50.000	52.679	-5.4	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.007	-6.0	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.791	-1.6	97	0.00
82 T	4-Chlorotoluene	50.000	53.327	-6.7	104	0.00
83 T	tert-Butylbenzene	50.000	53.079	-6.2	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.135	-6.3	103	0.00
85 T	sec-Butylbenzene	50.000	53.437	-6.9	105	0.00
86 T	p-Isopropyltoluene	50.000	55.097	-10.2	106	0.00
87 T	1,3-Dichlorobenzene	50.000	52.745	-5.5	105	0.00
88 T	1,4-Dichlorobenzene	50.000	51.453	-2.9	106	0.00
89 T	n-Butylbenzene	50.000	54.834	-9.7	108	0.00
90 T	Hexachloroethane	50.000	48.234	3.5	101	0.00
91 T	1,2-Dichlorobenzene	50.000	51.120	-2.2	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	43.947	12.1	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.726	-5.5	104	0.00
94 T	Hexachlorobutadiene	50.000	52.336	-4.7	108	0.00
95 T	Naphthalene	50.000	52.624	-5.2	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072821\
Data File : VX023490.D
Acq On : 28 Jul 2021 14:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 28 19:44:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.549	-3.1	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6