

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030274.D
 Acq On : 28 Jul 2022 16:58
 Operator : JC/MD
 Sample : N3839-08DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANKDL

Quant Time: Jul 29 05:58:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	249644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	144591	51.801	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.600%
35) Dibromofluoromethane	5.391	113	112301	47.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	8.653	98	420798	45.531	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.060%
62) 4-Bromofluorobenzene	11.085	95	150927	44.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.000%
Target Compounds						Qvalue
16) Acetone	2.386	43	2969	2.078	ug/l	93
17) Carbon Disulfide	2.514	76	129880	22.178	ug/l	99
20) Methylene Chloride	2.794	84	8669	1.057	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
Data File : VX030274.D
Acq On : 28 Jul 2022 16:58
Operator : JC/MD
Sample : N3839-08DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANKDL

Quant Time: Jul 29 05:58:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

