

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030278.D
 Acq On : 28 Jul 2022 18:32
 Operator : JC/MD
 Sample : N3773-16MEDL 10X
 Misc : 5.73g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DL-3(5-5.5)ME

Quant Time: Jul 29 05:59:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158809	53.672	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.340%	
35) Dibromofluoromethane	5.391	113	127165	51.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.340%	
50) Toluene-d8	8.653	98	487344	51.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.280%	
62) 4-Bromofluorobenzene	11.085	95	180656	51.077	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.160%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.501	56	8665	2.017	ug/l #	57
39) Methylcyclohexane	7.391	83	10118	2.213	ug/l	95
78) n-propylbenzene	11.311	91	20233	1.392	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	24337	2.387	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	45038	4.428	ug/l	99
95) Naphthalene	13.780	128	16233	1.314	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
Data File : VX030278.D
Acq On : 28 Jul 2022 18:32
Operator : JC/MD
Sample : N3773-16MEDL 10X
Misc : 5.73g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DL-3(5-5.5)ME

Quant Time: Jul 29 05:59:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

