

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030280.D
 Acq On : 28 Jul 2022 19:19
 Operator : JC/MD
 Sample : N3914-01ME 10X
 Misc : 7.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 METRO-DISPME

Quant Time: Jul 29 05:59:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	256564	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152707	52.386	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.780%			
35) Dibromofluoromethane	5.391	113	121714	49.626	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.260%			
50) Toluene-d8	8.653	98	475085	50.019	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.040%			
62) 4-Bromofluorobenzene	11.085	95	188349	53.432	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.860%			
Target Compounds					Qvalue	
40) Benzene	6.044	78	14758	1.329	ug/l	99
67) Ethyl Benzene	10.195	91	378623	32.347	ug/l	98
68) m/p-Xylenes	10.305	106	40879	8.429	ug/l	95
69) o-Xylene	10.646	106	89661	18.757	ug/l	93
73) Isopropylbenzene	10.963	105	90624	6.743	ug/l	97
78) n-propylbenzene	11.305	91	36218	2.258	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	151779	13.486	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	318633	28.378	ug/l	99
86) p-Isopropyltoluene	12.012	119	75835	6.639	ug/l	96
95) Naphthalene	13.786	128	13014688	954.085	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
Data File : VX030280.D
Acq On : 28 Jul 2022 19:19
Operator : JC/MD
Sample : N3914-01ME 10X
Misc : 7.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
METRO-DISPME

Quant Time: Jul 29 05:59:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

